

3/25/2018

Mr. Curt Lueckenhoff  
Missouri Dept. of Natural Resources  
2710 West Main

Jefferson City MO 65109

Project Name: Bridgeton Landfill

Project #:

Workorder #: 1803247

Dear Mr. Curt Lueckenhoff

The following report includes the data for the above referenced project for sample(s) received on 3/14/2018 at Air Toxics Ltd.

The data and associated QC analyzed by Modified TO-15 are compliant with the project requirements or laboratory criteria with the exception of the deviations noted in the attached case narrative.

Thank you for choosing Eurofins Air Toxics Inc. for your air analysis needs. Eurofins Air Toxics Inc. is committed to providing accurate data of the highest quality. Please feel free to contact the Project Manager: Brian Whittaker at 916-985-1000 if you have any questions regarding the data in this report.

Regards,



Brian Whittaker  
Project Manager

**WORK ORDER #: 1803247**

Work Order Summary

<b>CLIENT:</b>	Mr. Curt Lueckenhoff Missouri Dept. of Natural Resources 2710 West Main Jefferson City, MO 65109	<b>BILL TO:</b>	Accounts Payable Missouri Dept. of Natural Resources P.O Box 176 Jefferson City, MO 65102
<b>PHONE:</b>	573-526-3363	<b>P.O. #</b>	3ESP180688
<b>FAX:</b>	573-526-3350	<b>PROJECT #</b>	Bridgeton Landfill
<b>DATE RECEIVED:</b>	03/14/2018	<b>CONTACT:</b>	Brian Whittaker
<b>DATE COMPLETED:</b>	03/25/2018		

<u>FRACTION #</u>	<u>NAME</u>	<u>TEST</u>	<u>RECEIPT VAC./PRES.</u>	<u>FINAL PRESSURE</u>
01A	D1 (180993)	Modified TO-15	4.7 "Hg	5.2 psi
02A	U1 (180994)	Modified TO-15	4.9 "Hg	4.8 psi
03A	Lab Blank	Modified TO-15	NA	NA
04A	CCV	Modified TO-15	NA	NA
05A	LCS	Modified TO-15	NA	NA
05AA	LCSD	Modified TO-15	NA	NA

CERTIFIED BY:   
 Technical Director

DATE: 03/25/18

Certification numbers: AZ Licensure AZ0775, NJ NELAP - CA016, NY NELAP - 11291,  
 TX NELAP - T104704434-16-11, UT NELAP CA0093332016-7, VA NELAP - 8113, WA NELAP - C935  
 Name of Accreditation Body: NELAP/ORELAP (Oregon Environmental Laboratory Accreditation Program)  
 Accreditation number: CA300005, Effective date: 10/18/2016, Expiration date: 10/17/2017.

Eurofins Air Toxics Inc.. certifies that the test results contained in this report meet all requirements of the NELAC standards

**LABORATORY NARRATIVE**  
**Modified TO-15**  
**Missouri Dept. of Natural Resources**  
**Workorder# 1803247**

Two 6 Liter Summa Canister (100% Certified) samples were received on March 14, 2018. The laboratory performed analysis via modified EPA Method TO-15 using GC/MS in the full scan mode.

This workorder was independently validated prior to submittal using 'USEPA National Functional Guidelines' as generally applied to the analysis of volatile organic compounds in air. A rules-based, logic driven, independent validation engine was employed to assess completeness, evaluate pass/fail of relevant project quality control requirements and verification of all quantified amounts.

Method modifications taken to run these samples are summarized in the table below. Specific project requirements may over-ride the ATL modifications.

<i>Requirement</i>	<i>TO-15</i>	<i>ATL Modifications</i>
Initial Calibration	</=30% RSD with 2 compounds allowed out to < 40% RSD	</=30% RSD with 4 compounds allowed out to < 40% RSD
Blank and standards	Zero Air	UHP Nitrogen provides a higher purity gas matrix than zero air

**Receiving Notes**

There were no receiving discrepancies.

**Analytical Notes**

All Quality Control Limit exceedances and affected sample results are noted by flags. Each flag is defined at the bottom of this Case Narrative and on each Sample Result Summary page. Target compound non-detects in the samples that are associated with high bias in QC analyses have not been flagged.

**Definition of Data Qualifying Flags**

Eight qualifiers may have been used on the data analysis sheets and indicates as follows:

B - Compound present in laboratory blank greater than reporting limit (background subtraction not performed).

J - Estimated value.

E - Exceeds instrument calibration range.

S - Saturated peak.

Q - Exceeds quality control limits.

U - Compound analyzed for but not detected above the reporting limit, LOD, or MDL value. See data page for project specific U-flag definition.

UJ- Non-detected compound associated with low bias in the CCV

N - The identification is based on presumptive evidence.

File extensions may have been used on the data analysis sheets and indicates as follows:

a-File was requantified

b-File was quantified by a second column and detector

r1-File was requantified for the purpose of reissue

**Summary of Detected Compounds**  
**MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN**

**Client Sample ID: D1 (180993)**

**Lab ID#: 1803247-01A**

<b>Compound</b>	<b>Rpt. Limit (ppbv)</b>	<b>Amount (ppbv)</b>	<b>Rpt. Limit (ug/m3)</b>	<b>Amount (ug/m3)</b>
Freon 12	0.16	0.42	0.79	2.1
Freon 11	0.16	0.21	0.90	1.2
Ethanol	0.80	0.92	1.5	1.7
Acetone	0.80	2.9	1.9	7.0

**Client Sample ID: U1 (180994)**

**Lab ID#: 1803247-02A**

<b>Compound</b>	<b>Rpt. Limit (ppbv)</b>	<b>Amount (ppbv)</b>	<b>Rpt. Limit (ug/m3)</b>	<b>Amount (ug/m3)</b>
Freon 12	0.16	0.44	0.78	2.2
Freon 11	0.16	0.21	0.89	1.2
Ethanol	0.79	1.0	1.5	1.9
Acetone	0.79	2.1	1.9	4.9



Air Toxics

Client Sample ID: D1 (180993)

Lab ID#: 1803247-01A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031507	Date of Collection:	3/8/18 1:15:00 PM
Dil. Factor:	1.60	Date of Analysis:	3/15/18 11:46 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.16	0.42	0.79	2.1
Freon 114	0.16	Not Detected	1.1	Not Detected
Chloromethane	0.80	Not Detected	1.6	Not Detected
Vinyl Chloride	0.16	Not Detected	0.41	Not Detected
1,3-Butadiene	0.16	Not Detected	0.35	Not Detected
Bromomethane	0.80	Not Detected	3.1	Not Detected
Chloroethane	0.80	Not Detected	2.1	Not Detected
Freon 11	0.16	0.21	0.90	1.2
Ethanol	0.80	0.92	1.5	1.7
Freon 113	0.16	Not Detected	1.2	Not Detected
1,1-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Acetone	0.80	2.9	1.9	7.0
2-Propanol	0.80	Not Detected	2.0	Not Detected
Carbon Disulfide	0.80	Not Detected	2.5	Not Detected
3-Chloropropene	0.80	Not Detected	2.5	Not Detected
Methylene Chloride	0.32	Not Detected	1.1	Not Detected
Methyl tert-butyl ether	0.16	Not Detected	0.58	Not Detected
trans-1,2-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Hexane	0.16	Not Detected	0.56	Not Detected
1,1-Dichloroethane	0.16	Not Detected	0.65	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.80	Not Detected	2.4	Not Detected
cis-1,2-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Tetrahydrofuran	0.80	Not Detected	2.4	Not Detected
Chloroform	0.16	Not Detected	0.78	Not Detected
1,1,1-Trichloroethane	0.16	Not Detected	0.87	Not Detected
Cyclohexane	0.16	Not Detected	0.55	Not Detected
Carbon Tetrachloride	0.16	Not Detected	1.0	Not Detected
2,2,4-Trimethylpentane	0.80	Not Detected	3.7	Not Detected
Benzene	0.16	Not Detected	0.51	Not Detected
1,2-Dichloroethane	0.16	Not Detected	0.65	Not Detected
Heptane	0.16	Not Detected	0.66	Not Detected
Trichloroethene	0.16	Not Detected	0.86	Not Detected
1,2-Dichloropropane	0.16	Not Detected	0.74	Not Detected
1,4-Dioxane	0.16	Not Detected	0.58	Not Detected
Bromodichloromethane	0.16	Not Detected	1.1	Not Detected
cis-1,3-Dichloropropene	0.16	Not Detected	0.73	Not Detected
4-Methyl-2-pentanone	0.16	Not Detected	0.66	Not Detected
Toluene	0.16	Not Detected	0.60	Not Detected
trans-1,3-Dichloropropene	0.16	Not Detected	0.73	Not Detected
1,1,2-Trichloroethane	0.16	Not Detected	0.87	Not Detected
Tetrachloroethene	0.16	Not Detected	1.1	Not Detected
2-Hexanone	0.80	Not Detected	3.3	Not Detected



Client Sample ID: D1 (180993)

Lab ID#: 1803247-01A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031507	Date of Collection:	3/8/18 1:15:00 PM
Dil. Factor:	1.60	Date of Analysis:	3/15/18 11:46 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.16	Not Detected	1.4	Not Detected
1,2-Dibromoethane (EDB)	0.16	Not Detected	1.2	Not Detected
Chlorobenzene	0.16	Not Detected	0.74	Not Detected
Ethyl Benzene	0.16	Not Detected	0.69	Not Detected
m,p-Xylene	0.16	Not Detected	0.69	Not Detected
o-Xylene	0.16	Not Detected	0.69	Not Detected
Styrene	0.16	Not Detected	0.68	Not Detected
Bromoform	0.16	Not Detected	1.6	Not Detected
Cumene	0.16	Not Detected	0.79	Not Detected
1,1,2,2-Tetrachloroethane	0.16	Not Detected	1.1	Not Detected
Propylbenzene	0.16	Not Detected	0.79	Not Detected
4-Ethyltoluene	0.16	Not Detected	0.79	Not Detected
1,3,5-Trimethylbenzene	0.16	Not Detected	0.79	Not Detected
1,2,4-Trimethylbenzene	0.16	Not Detected	0.79	Not Detected
1,3-Dichlorobenzene	0.16	Not Detected	0.96	Not Detected
1,4-Dichlorobenzene	0.16	Not Detected	0.96	Not Detected
alpha-Chlorotoluene	0.16	Not Detected	0.83	Not Detected
1,2-Dichlorobenzene	0.16	Not Detected	0.96	Not Detected
1,2,4-Trichlorobenzene	0.80	Not Detected UJ	5.9	Not Detected UJ
Hexachlorobutadiene	0.80	Not Detected	8.5	Not Detected

UJ = Analyte associated with low bias in the CCV.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	118	70-130
Toluene-d8	99	70-130
4-Bromofluorobenzene	93	70-130



Air Toxics

Client Sample ID: U1 (180994)

Lab ID#: 1803247-02A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031508	Date of Collection:	3/8/18 1:00:00 PM
Dil. Factor:	1.58	Date of Analysis:	3/15/18 12:20 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.16	0.44	0.78	2.2
Freon 114	0.16	Not Detected	1.1	Not Detected
Chloromethane	0.79	Not Detected	1.6	Not Detected
Vinyl Chloride	0.16	Not Detected	0.40	Not Detected
1,3-Butadiene	0.16	Not Detected	0.35	Not Detected
Bromomethane	0.79	Not Detected	3.1	Not Detected
Chloroethane	0.79	Not Detected	2.1	Not Detected
Freon 11	0.16	0.21	0.89	1.2
Ethanol	0.79	1.0	1.5	1.9
Freon 113	0.16	Not Detected	1.2	Not Detected
1,1-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Acetone	0.79	2.1	1.9	4.9
2-Propanol	0.79	Not Detected	1.9	Not Detected
Carbon Disulfide	0.79	Not Detected	2.5	Not Detected
3-Chloropropene	0.79	Not Detected	2.5	Not Detected
Methylene Chloride	0.32	Not Detected	1.1	Not Detected
Methyl tert-butyl ether	0.16	Not Detected	0.57	Not Detected
trans-1,2-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Hexane	0.16	Not Detected	0.56	Not Detected
1,1-Dichloroethane	0.16	Not Detected	0.64	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.79	Not Detected	2.3	Not Detected
cis-1,2-Dichloroethene	0.16	Not Detected	0.63	Not Detected
Tetrahydrofuran	0.79	Not Detected	2.3	Not Detected
Chloroform	0.16	Not Detected	0.77	Not Detected
1,1,1-Trichloroethane	0.16	Not Detected	0.86	Not Detected
Cyclohexane	0.16	Not Detected	0.54	Not Detected
Carbon Tetrachloride	0.16	Not Detected	0.99	Not Detected
2,2,4-Trimethylpentane	0.79	Not Detected	3.7	Not Detected
Benzene	0.16	Not Detected	0.50	Not Detected
1,2-Dichloroethane	0.16	Not Detected	0.64	Not Detected
Heptane	0.16	Not Detected	0.65	Not Detected
Trichloroethene	0.16	Not Detected	0.85	Not Detected
1,2-Dichloropropane	0.16	Not Detected	0.73	Not Detected
1,4-Dioxane	0.16	Not Detected	0.57	Not Detected
Bromodichloromethane	0.16	Not Detected	1.0	Not Detected
cis-1,3-Dichloropropene	0.16	Not Detected	0.72	Not Detected
4-Methyl-2-pentanone	0.16	Not Detected	0.65	Not Detected
Toluene	0.16	Not Detected	0.60	Not Detected
trans-1,3-Dichloropropene	0.16	Not Detected	0.72	Not Detected
1,1,2-Trichloroethane	0.16	Not Detected	0.86	Not Detected
Tetrachloroethene	0.16	Not Detected	1.1	Not Detected
2-Hexanone	0.79	Not Detected	3.2	Not Detected





Client Sample ID: U1 (180994)

Lab ID#: 1803247-02A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031508	Date of Collection:	3/8/18 1:00:00 PM
Dil. Factor:	1.58	Date of Analysis:	3/15/18 12:20 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.16	Not Detected	1.3	Not Detected
1,2-Dibromoethane (EDB)	0.16	Not Detected	1.2	Not Detected
Chlorobenzene	0.16	Not Detected	0.73	Not Detected
Ethyl Benzene	0.16	Not Detected	0.69	Not Detected
m,p-Xylene	0.16	Not Detected	0.69	Not Detected
o-Xylene	0.16	Not Detected	0.69	Not Detected
Styrene	0.16	Not Detected	0.67	Not Detected
Bromoform	0.16	Not Detected	1.6	Not Detected
Cumene	0.16	Not Detected	0.78	Not Detected
1,1,2,2-Tetrachloroethane	0.16	Not Detected	1.1	Not Detected
Propylbenzene	0.16	Not Detected	0.78	Not Detected
4-Ethyltoluene	0.16	Not Detected	0.78	Not Detected
1,3,5-Trimethylbenzene	0.16	Not Detected	0.78	Not Detected
1,2,4-Trimethylbenzene	0.16	Not Detected	0.78	Not Detected
1,3-Dichlorobenzene	0.16	Not Detected	0.95	Not Detected
1,4-Dichlorobenzene	0.16	Not Detected	0.95	Not Detected
alpha-Chlorotoluene	0.16	Not Detected	0.82	Not Detected
1,2-Dichlorobenzene	0.16	Not Detected	0.95	Not Detected
1,2,4-Trichlorobenzene	0.79	Not Detected UJ	5.9	Not Detected UJ
Hexachlorobutadiene	0.79	Not Detected	8.4	Not Detected

UJ = Analyte associated with low bias in the CCV.

Container Type: 6 Liter Summa Canister (100% Certified)

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	116	70-130
Toluene-d8	97	70-130
4-Bromofluorobenzene	93	70-130



Air Toxics

Client Sample ID: Lab Blank

Lab ID#: 1803247-03A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031506	Date of Collection:	NA
Dil. Factor:	1.00	Date of Analysis:	3/15/18 11:00 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Freon 12	0.10	Not Detected	0.49	Not Detected
Freon 114	0.10	Not Detected	0.70	Not Detected
Chloromethane	0.50	Not Detected	1.0	Not Detected
Vinyl Chloride	0.10	Not Detected	0.26	Not Detected
1,3-Butadiene	0.10	Not Detected	0.22	Not Detected
Bromomethane	0.50	Not Detected	1.9	Not Detected
Chloroethane	0.50	Not Detected	1.3	Not Detected
Freon 11	0.10	Not Detected	0.56	Not Detected
Ethanol	0.50	Not Detected	0.94	Not Detected
Freon 113	0.10	Not Detected	0.77	Not Detected
1,1-Dichloroethene	0.10	Not Detected	0.40	Not Detected
Acetone	0.50	Not Detected	1.2	Not Detected
2-Propanol	0.50	Not Detected	1.2	Not Detected
Carbon Disulfide	0.50	Not Detected	1.6	Not Detected
3-Chloropropene	0.50	Not Detected	1.6	Not Detected
Methylene Chloride	0.20	Not Detected	0.69	Not Detected
Methyl tert-butyl ether	0.10	Not Detected	0.36	Not Detected
trans-1,2-Dichloroethene	0.10	Not Detected	0.40	Not Detected
Hexane	0.10	Not Detected	0.35	Not Detected
1,1-Dichloroethane	0.10	Not Detected	0.40	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.50	Not Detected	1.5	Not Detected
cis-1,2-Dichloroethene	0.10	Not Detected	0.40	Not Detected
Tetrahydrofuran	0.50	Not Detected	1.5	Not Detected
Chloroform	0.10	Not Detected	0.49	Not Detected
1,1,1-Trichloroethane	0.10	Not Detected	0.54	Not Detected
Cyclohexane	0.10	Not Detected	0.34	Not Detected
Carbon Tetrachloride	0.10	Not Detected	0.63	Not Detected
2,2,4-Trimethylpentane	0.50	Not Detected	2.3	Not Detected
Benzene	0.10	Not Detected	0.32	Not Detected
1,2-Dichloroethane	0.10	Not Detected	0.40	Not Detected
Heptane	0.10	Not Detected	0.41	Not Detected
Trichloroethene	0.10	Not Detected	0.54	Not Detected
1,2-Dichloropropane	0.10	Not Detected	0.46	Not Detected
1,4-Dioxane	0.10	Not Detected	0.36	Not Detected
Bromodichloromethane	0.10	Not Detected	0.67	Not Detected
cis-1,3-Dichloropropene	0.10	Not Detected	0.45	Not Detected
4-Methyl-2-pentanone	0.10	Not Detected	0.41	Not Detected
Toluene	0.10	Not Detected	0.38	Not Detected
trans-1,3-Dichloropropene	0.10	Not Detected	0.45	Not Detected
1,1,2-Trichloroethane	0.10	Not Detected	0.54	Not Detected
Tetrachloroethene	0.10	Not Detected	0.68	Not Detected
2-Hexanone	0.50	Not Detected	2.0	Not Detected



Client Sample ID: Lab Blank

Lab ID#: 1803247-03A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031506	Date of Collection:	NA
Dil. Factor:	1.00	Date of Analysis:	3/15/18 11:00 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Dibromochloromethane	0.10	Not Detected	0.85	Not Detected
1,2-Dibromoethane (EDB)	0.10	Not Detected	0.77	Not Detected
Chlorobenzene	0.10	Not Detected	0.46	Not Detected
Ethyl Benzene	0.10	Not Detected	0.43	Not Detected
m,p-Xylene	0.10	Not Detected	0.43	Not Detected
o-Xylene	0.10	Not Detected	0.43	Not Detected
Styrene	0.10	Not Detected	0.42	Not Detected
Bromoform	0.10	Not Detected	1.0	Not Detected
Cumene	0.10	Not Detected	0.49	Not Detected
1,1,2,2-Tetrachloroethane	0.10	Not Detected	0.69	Not Detected
Propylbenzene	0.10	Not Detected	0.49	Not Detected
4-Ethyltoluene	0.10	Not Detected	0.49	Not Detected
1,3,5-Trimethylbenzene	0.10	Not Detected	0.49	Not Detected
1,2,4-Trimethylbenzene	0.10	Not Detected	0.49	Not Detected
1,3-Dichlorobenzene	0.10	Not Detected	0.60	Not Detected
1,4-Dichlorobenzene	0.10	Not Detected	0.60	Not Detected
alpha-Chlorotoluene	0.10	Not Detected	0.52	Not Detected
1,2-Dichlorobenzene	0.10	Not Detected	0.60	Not Detected
1,2,4-Trichlorobenzene	0.50	Not Detected UJ	3.7	Not Detected UJ
Hexachlorobutadiene	0.50	Not Detected	5.3	Not Detected

UJ = Analyte associated with low bias in the CCV.

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	113	70-130
Toluene-d8	96	70-130
4-Bromofluorobenzene	112	70-130



Air Toxics

Client Sample ID: CCV

Lab ID#: 1803247-04A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031503	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 3/15/18 09:14 AM

Compound	%Recovery
Freon 12	104
Freon 114	97
Chloromethane	97
Vinyl Chloride	98
1,3-Butadiene	95
Bromomethane	104
Chloroethane	100
Freon 11	98
Ethanol	104
Freon 113	98
1,1-Dichloroethene	99
Acetone	100
2-Propanol	95
Carbon Disulfide	98
3-Chloropropene	100
Methylene Chloride	97
Methyl tert-butyl ether	104
trans-1,2-Dichloroethene	91
Hexane	102
1,1-Dichloroethane	98
2-Butanone (Methyl Ethyl Ketone)	104
cis-1,2-Dichloroethene	104
Tetrahydrofuran	102
Chloroform	99
1,1,1-Trichloroethane	99
Cyclohexane	104
Carbon Tetrachloride	106
2,2,4-Trimethylpentane	114
Benzene	102
1,2-Dichloroethane	100
Heptane	110
Trichloroethene	100
1,2-Dichloropropane	102
1,4-Dioxane	107
Bromodichloromethane	105
cis-1,3-Dichloropropene	102
4-Methyl-2-pentanone	114
Toluene	104
trans-1,3-Dichloropropene	103
1,1,2-Trichloroethane	101
Tetrachloroethene	99
2-Hexanone	112



Air Toxics

Client Sample ID: CCV

Lab ID#: 1803247-04A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031503	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 3/15/18 09:14 AM

Compound	%Recovery
Dibromochloromethane	109
1,2-Dibromoethane (EDB)	106
Chlorobenzene	101
Ethyl Benzene	111
m,p-Xylene	110
o-Xylene	115
Styrene	117
Bromoform	111
Cumene	104
1,1,2,2-Tetrachloroethane	106
Propylbenzene	102
4-Ethyltoluene	112
1,3,5-Trimethylbenzene	114
1,2,4-Trimethylbenzene	111
1,3-Dichlorobenzene	98
1,4-Dichlorobenzene	97
alpha-Chlorotoluene	100
1,2-Dichlorobenzene	97
1,2,4-Trichlorobenzene	65 Q
Hexachlorobutadiene	70

Q = Exceeds Quality Control limits.

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	98	70-130
Toluene-d8	102	70-130
4-Bromofluorobenzene	97	70-130



Air Toxics

Client Sample ID: LCS

Lab ID#: 1803247-05A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031504	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 3/15/18 09:55 AM

Compound	%Recovery	Method Limits
Freon 12	93	70-130
Freon 114	91	70-130
Chloromethane	84	70-130
Vinyl Chloride	94	70-130
1,3-Butadiene	89	70-130
Bromomethane	100	70-130
Chloroethane	95	70-130
Freon 11	89	70-130
Ethanol	91	70-130
Freon 113	85	70-130
1,1-Dichloroethene	90	70-130
Acetone	90	70-130
2-Propanol	94	70-130
Carbon Disulfide	80	70-130
3-Chloropropene	89	70-130
Methylene Chloride	89	70-130
Methyl tert-butyl ether	91	70-130
trans-1,2-Dichloroethene	76	70-130
Hexane	92	70-130
1,1-Dichloroethane	89	70-130
2-Butanone (Methyl Ethyl Ketone)	93	70-130
cis-1,2-Dichloroethene	99	70-130
Tetrahydrofuran	93	70-130
Chloroform	88	70-130
1,1,1-Trichloroethane	88	70-130
Cyclohexane	93	70-130
Carbon Tetrachloride	92	70-130
2,2,4-Trimethylpentane	103	70-130
Benzene	91	70-130
1,2-Dichloroethane	89	70-130
Heptane	98	70-130
Trichloroethene	90	70-130
1,2-Dichloropropane	93	70-130
1,4-Dioxane	91	70-130
Bromodichloromethane	99	70-130
cis-1,3-Dichloropropene	92	70-130
4-Methyl-2-pentanone	102	70-130
Toluene	93	70-130
trans-1,3-Dichloropropene	93	70-130
1,1,2-Trichloroethane	94	70-130
Tetrachloroethene	89	70-130
2-Hexanone	97	70-130

Client Sample ID: LCS

Lab ID#: 1803247-05A

**MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN**

File Name:	21031504	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 3/15/18 09:55 AM

Compound	%Recovery	Method Limits
Dibromochloromethane	101	70-130
1,2-Dibromoethane (EDB)	96	70-130
Chlorobenzene	93	70-130
Ethyl Benzene	101	70-130
m,p-Xylene	101	70-130
o-Xylene	106	70-130
Styrene	102	70-130
Bromoform	109	70-130
Cumene	94	70-130
1,1,2,2-Tetrachloroethane	103	70-130
Propylbenzene	93	70-130
4-Ethyltoluene	101	70-130
1,3,5-Trimethylbenzene	105	70-130
1,2,4-Trimethylbenzene	104	70-130
1,3-Dichlorobenzene	92	70-130
1,4-Dichlorobenzene	88	70-130
alpha-Chlorotoluene	101	70-130
1,2-Dichlorobenzene	92	70-130
1,2,4-Trichlorobenzene	67 Q	70-130
Hexachlorobutadiene	75	70-130

Q = Exceeds Quality Control limits.

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	101	70-130
Toluene-d8	101	70-130
4-Bromofluorobenzene	94	70-130



Client Sample ID: LCSD

Lab ID#: 1803247-05AA

## MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	21031505	Date of Collection:	NA
Dil. Factor:	1.00	Date of Analysis:	3/15/18 10:28 AM

Compound	%Recovery	Method Limits
Freon 12	92	70-130
Freon 114	91	70-130
Chloromethane	84	70-130
Vinyl Chloride	93	70-130
1,3-Butadiene	89	70-130
Bromomethane	98	70-130
Chloroethane	94	70-130
Freon 11	90	70-130
Ethanol	91	70-130
Freon 113	84	70-130
1,1-Dichloroethene	90	70-130
Acetone	89	70-130
2-Propanol	94	70-130
Carbon Disulfide	79	70-130
3-Chloropropene	89	70-130
Methylene Chloride	88	70-130
Methyl tert-butyl ether	92	70-130
trans-1,2-Dichloroethene	75	70-130
Hexane	91	70-130
1,1-Dichloroethane	89	70-130
2-Butanone (Methyl Ethyl Ketone)	93	70-130
cis-1,2-Dichloroethene	98	70-130
Tetrahydrofuran	92	70-130
Chloroform	88	70-130
1,1,1-Trichloroethane	88	70-130
Cyclohexane	93	70-130
Carbon Tetrachloride	92	70-130
2,2,4-Trimethylpentane	101	70-130
Benzene	89	70-130
1,2-Dichloroethane	89	70-130
Heptane	97	70-130
Trichloroethene	90	70-130
1,2-Dichloropropane	92	70-130
1,4-Dioxane	91	70-130
Bromodichloromethane	101	70-130
cis-1,3-Dichloropropene	92	70-130
4-Methyl-2-pentanone	102	70-130
Toluene	92	70-130
trans-1,3-Dichloropropene	96	70-130
1,1,2-Trichloroethane	96	70-130
Tetrachloroethene	92	70-130
2-Hexanone	100	70-130



Client Sample ID: LCSD

Lab ID#: 1803247-05AA

**MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN**

<b>File Name:</b>	<b>21031505</b>	<b>Date of Collection: NA</b>
<b>Dil. Factor:</b>	<b>1.00</b>	<b>Date of Analysis: 3/15/18 10:28 AM</b>

Compound	%Recovery	Method Limits
Dibromochloromethane	103	70-130
1,2-Dibromoethane (EDB)	99	70-130
Chlorobenzene	95	70-130
Ethyl Benzene	102	70-130
m,p-Xylene	101	70-130
o-Xylene	107	70-130
Styrene	106	70-130
Bromoform	112	70-130
Cumene	98	70-130
1,1,2,2-Tetrachloroethane	107	70-130
Propylbenzene	99	70-130
4-Ethyltoluene	107	70-130
1,3,5-Trimethylbenzene	110	70-130
1,2,4-Trimethylbenzene	108	70-130
1,3-Dichlorobenzene	92	70-130
1,4-Dichlorobenzene	89	70-130
alpha-Chlorotoluene	104	70-130
1,2-Dichlorobenzene	92	70-130
1,2,4-Trichlorobenzene	66 Q	70-130
Hexachlorobutadiene	74	70-130

Q = Exceeds Quality Control limits.

**Container Type: NA - Not Applicable**

Surrogates	%Recovery	Method Limits
1,2-Dichloroethane-d4	99	70-130
Toluene-d8	100	70-130
4-Bromofluorobenzene	99	70-130